Spares, consumables and installation

Consumables for Purelab 1 and 3

Cat. No.	Code	Description
WP/44240	LC241	Pre-treatment pack
WP/44242	LC240	Reverse Osmosis module, 10 litres an hour

** It is recommended that the pre-treatment pack is changed every 6 months and the Reverse Osmosis module every 2- 3 years

Optional consumable

Cat. No.	Code	Description
WP/44250	LC181	E-Cartridge, requires degassing kit
WP/44252	LA775	Degassing kit for Chorus 3 (RO) and 2 (RO/DI)

**It is recommended that the E-cartridge is changed every 3 years, a degassing kit is required to do this

Installation kits and accessories

Cat. No.	Code	Description
WP/44260	LA762	Installation kit, basic, tubing and bends
WP/44262	LA779	Installation kit, remote mount
WP/44264	LA795	Mounting kit, wall, integrated
WP/44266	LA776	Tool kit
WP/43970	LA512	Pressure regulator
WP/44268	LC233	Bypass block
WP/44270	LA732	Foot switch

Notes